

**Institute of Public Health in Ostrava**

Centre of Hygienic Laboratories

CAI Accredited Testing Laboratory No. 1393 according to SN EN ISO/IEC 17025  
Partyzánské nám stí 7, 702 00 Ostrava**TEST REPORT No. 1440/2016****Customer :** CRYSTAL COLLOIDALS  
Thomas Alva Edisonweg 3  
6045 GN Roermond  
NL**Set No. :** 1258  
**Sample Received :** 13.1.2016 15:00  
**Sample Analyzed :** 13.1.2016 - 20.1.2016  
**Ref. No. :** ZU/ZU/01044/2016  
**File No. :** S-ZU/ZU/01044/2016  
**File code :** 4.0.3

<b>Sample No. :</b>	<b>4089</b>
<b>Sampling date :</b>	Not mentioned
<b>Sampling time:</b>	Not mentioned
<b>Sample name:</b>	Crystal colloids Lot.nr: Ag011015
<b>Expiry date :</b>	2019
<b>Sample Type :</b>	Drinking water
<b>Mode of sampling :</b>	Not mentioned

**Results - chemical analysis**

Parameter	Value	Unit	Type	Method used	Uncertainty
Ag	10,7	mg/l	A	SOP OV 201 <sup>2</sup>	±20%
Size Distribution by Number	See the Annex to the Report		N	SOP OV 2039 <sup>2</sup>	-

**Notice to sampling :** The sampling itself is not a subject of accreditation.**Method specification :**

SOP OV 201 ( SN EN ISO 17294-1, SN EN ISO 17294-2)

**Laboratory workplace :**<sup>(2)</sup> - Analyses performed at Ostrava (Partyzánské nám. 7, 702 00 Ostrava)

Methods in TYPE column: "A" accredited test,"N" non-accredited test

&lt; - the result is below the detection limit, &gt; - the result is higher than the value presented

Results deal with tested samples only.

Without a written consent of the laboratory, this Report can be reproduced only complete.

These expanded uncertainties of measurement are obtained by multiplying of standard uncertainty of measurement by extending coefficient k=2 (for confidence level 95%). Uncertainty of sampling not included.

**Head of Hygienic Laboratories Center :** Došká ová Šárka, RNDr.**Checked by :** Lach Karel, Ing. CSc.**Completed by :** Lach Karel, Ing. CSc.**Number of pages :** 1**Date :** 21.1.2016Ing. Vladimíra N mcová  
Head of Department of Inorganic Analyses