

Institute of Public Health in Ostrava
Centre of Hygienic Laboratories
Partyzánské nám stí 7, 702 00 Ostrava

TEST REPORT No. 68937/2016

Customer : CRYSTAL COLLOIDALS
Thomas Alva Edisonweg 3
6045 GN Roermond
NL

Set No. : 42556
Sample Received : 25.11.2016 14:00
Sample Analyzed : 25.11.2016 - 6.12.2016
Ref. No. : ZU/ZU/01044/2016
File No. : S-ZU/ZU/01044/2016
File code : 4.0.3

Sample No. : 131565
Sampling date : 23.11.2016 **Sampling time:**Not mentioned
Sample name: Crystal colloids Lot.nr: Au 191016
Sample Type : Drinking water
Sampled by : Customer
Mode of sampling : Not mentioned

Results - chemical analysis

Parameter	Value	Unit	Type	Method used	Uncertainty
Au	10,1	mg/l	N	SOP OV 201 ²	±20%

Notice to sampling : The sampling itself is not a subject of accreditation.

Method specification :

SOP OV 201 (SN EN ISO 17294-1, SN EN ISO 17294-2)

Laboratory workplace :

⁽²⁾ - Analyses performed at Ostrava (Partyzánské nám. 7, 702 00 Ostrava)

Methods in TYPE column: "N" non-accredited test

< - the result is below the detection limit, > - the result is higher than the value presented

Results deal with tested samples only.

Without a written consent of the laboratory, this Report can be reproduced only complete.

These expanded uncertainties of measurement are obtained by multiplying of standard uncertainty of measurement by extending coefficient k=2 (for confidence level 95%). Uncertainty of sampling not included.

Head of Hygienic Laboratories Center : Došká ová Šárka, RNDr.

Checked by : Lach Karel, Ing. CSc.

Completed by : Lach Karel, Ing. CSc.

Number of pages : 1

Date : 6.12.2016

Ing. Vladimíra N mcová
Head of Department of Inorganic Analyses