

**Institute of Public Health in Ostrava**

Centre of Hygienic Laboratories

CAI Accredited Testing Laboratory No.1393 according to SN EN ISO/IEC 17025:2005  
Partyzánské nám. stí 7, 702 00 Ostrava**TEST REPORT No. 13582/2017****Customer :** CRYSTAL COLLOIDALS  
Thomas Alva Edisonweg 3  
6045 GN Roermond  
NL**Set No. :** 9239  
**Sample Received :** 27.3.2017 7:00  
**Sample Analyzed :** 27.3.2017 - 27.3.2017  
**Ref. No. :** ZU/ZU/01044/2016  
**File No. :** S-ZU/ZU/01044/2016  
**File code :** 4.0.3

<b>Sample No. :</b>	<b>29725</b>
<b>Sampling date :</b>	Not mentioned
<b>Sampling time:</b>	Not mentioned
<b>Sample name:</b>	Crystal colloids Lot.nr: Cu220317/ Expiration date:10-17
<b>Sample Type :</b>	Drinking water
<b>Sampled by :</b>	Customer
<b>Mode of sampling :</b>	Not mentioned

**Results - chemical analysis**

Parameter	Value	Unit	Type	Method used	Uncertainty
Cu	11,3	mg/l	A	SOP OV 201 <sup>2</sup>	±20%

**Notice to sampling :** The sampling itself is not a subject of accreditation.**Method specification :**

SOP OV 201 ( SN EN ISO 17294-1, SN EN ISO 17294-2)

**Laboratory workplace :**<sup>(2)</sup> - Analyses performed at Ostrava (Partyzánské nám. 7, 702 00 Ostrava)

Methods in TYPE column: "A" accredited test

&lt; - the result is below the detection limit, &gt; - the result is higher than the value presented

Results deal with tested samples only.

Without a written consent of the laboratory, this Report can be reproduced only complete.

These expanded uncertainties of measurement are obtained by multiplying of standard uncertainty of measurement by extending coefficient k=2 (for confidence level 95%). Uncertainty of sampling not included.

**Head of Hygienic Laboratories Center :** Došká ová Šárka, RNDr.**Checked by :** Lach Karel, Ing. CSc.**Completed by :** Lach Karel, Ing. CSc.**Number of pages :** 1**Date :** 27.3.2017Ing. Vladimíra N mcová  
Head of Department of Inorganic Analyses