

Institute of Public Health in Ostrava  
Centre of Hygienic Laboratories  
Partyzánské nám stí 7, 702 00 Ostrava

**TEST REPORT No. 55458/2016**

**Customer :** CRYSTAL COLLOIDALS  
Thomas Alva Edisonweg 3  
6045 GN Roermond  
NL

**Set No. :** 35520  
**Sample Received :** 6.10.2016 15:00  
**Sample Analyzed :** 6.10.2016 - 7.10.2016  
**Ref. No. :** ZU/ZU/01044/2016  
**File No. :** S-ZU/ZU/01044/2016  
**File code :** 4.0.3

<b>Sample No. :</b>	<b>110269</b>	<b>Sampling time:</b>	Not mentioned
<b>Sampling date :</b>	3.10.2016		
<b>Sample name:</b>	Crystal colloids Pt31016		
<b>Sample Type :</b>	Drinking water		
<b>Sampled by :</b>	Customer		
<b>Mode of sampling :</b>	Not mentioned		
<b>Purpose :</b>	on the request customers		

**Results - chemical analysis**

Parameter	Value	Unit	Type	Method used	Uncertainty
Pt	5,14	mg/l	N	SOP OV 201 <sup>2</sup>	±20%

**Notice to sampling :** The sampling itself is not a subject of accreditation.

**Method specification :**

SOP OV 201 ( SN EN ISO 17294-1, SN EN ISO 17294-2)

**Laboratory workplace :**

<sup>(2)</sup> - Analyses performed at Ostrava (Partyzánské nám. 7, 702 00 Ostrava)

Methods in TYPE column: "N" non-accredited test

< - the result is below the detection limit, > - the result is higher than the value presented

Results deal with tested samples only.

Without a written consent of the laboratory, this Report can be reproduced only complete.

These expanded uncertainties of measurement are obtained by multiplying of standard uncertainty of measurement by extending coefficient k=2 (for confidence level 95%). Uncertainty of sampling not included.

**Head of Hygienic Laboratories Center :** Došká ová Šárka, RNDr.

**Checked by :** Lach Karel, Ing. CSc.

**Completed by :** Lach Karel, Ing. CSc.

**Number of pages :** 1

**Date :** 7.10.2016

Ing. Vladimíra N mcová  
Head of Department of Inorganic Analyses