

**Institute of Public Health in Ostrava**  
**Centre of Hygienic Laboratories**  
**Partyzánské nám stí 7, 702 00 Ostrava**

**TEST REPORT No. 20877/2017**

**Customer :** CRYSTAL COLLOIDALS  
Thomas Alva Edisonweg 3  
6045 GN Roermond  
NL

**Set No. :** 13234  
**Sample Received :** 26.4.2017 11:00  
**Sample Analyzed :** 26.4.2017 - 28.4.2017  
**Ref. No. :** ZU/ZU/01044/2016  
**File No. :** S-ZU/ZU/01044/2016  
**File code :** 4.0.3

**Sample No. :** 42702  
**Sampling date :** Not mentioned **Sampling time:**Not mentioned  
**Sample name:** Crystal colloids Lot.nr: Pt200417/ Expiration date:2021  
**Sample Type :** Drinking water  
**Sampled by :** Customer  
**Mode of sampling :** Not mentioned

**Results - chemical analysis**

Parameter	Value	Unit	Type	Method used	Uncertainty
Pt	5,09	mg/l	N	SOP OV 201 <sup>2</sup>	±20%

**Notice to sampling :** The sampling itself is not a subject of accreditation.

**Method specification :**

SOP OV 201 ( SN EN ISO 17294-1, SN EN ISO 17294-2)

**Laboratory workplace :**

<sup>(2)</sup> - Analyses performed at Ostrava (Partyzánské nám. 7, 702 00 Ostrava)

Methods in TYPE column: "N" non-accredited test

< - the result is below the detection limit, > - the result is higher than the value presented

Results deal with tested samples only.

Without a written consent of the laboratory, this Report can be reproduced only complete.

These expanded uncertainties of measurement are obtained by multiplying of standard uncertainty of measurement by extending coefficient k=2 (for confidence level 95%). Uncertainty of sampling not included.

**Head of Hygienic Laboratories Center :** Došká ová Šárka, RNDr.

**Checked by :** Lach Karel, Ing. CSc.

**Completed by :** Lach Karel, Ing. CSc.

**Number of pages :** 1

**Date :** 28.4.2017

Ing. Vladimíra N mcová  
Head of Department of Inorganic Analyses