

**Institute of Public Health in Ostrava**

Centre of Hygienic Laboratories

CAI Accredited Testing Laboratory No.1393 according to SN EN ISO/IEC 17025:2005

Partyzánské nám. stí 7, 702 00 Ostrava

TEST REPORT No. 57301/2016**Customer :** CRYSTAL COLLOIDALS

Thomas Alva Edisonweg 3

6045 GN Roermond

NL

Set No. : 35663**Sample Received :** 7.10.2016 15:00**Sample Analyzed :** 7.10.2016 - 17.10.2016**Ref. No. :** ZU/ZU/01044/2016**File No. :** S-ZU/ZU/01044/2016**File code :** 4.0.3

| | |
|---------------------------|-----------------------------------|
| Sample No. : | 110725 |
| Sampling date : | 5.10.2016 |
| Sampling time : | Not mentioned |
| Sample name : | Crystal colloids Lot.nr: Ag031016 |
| Sample Type : | Drinking water |
| Sampled by : | Customer |
| Mode of sampling : | Not mentioned |
| Purpose : | on the request customers |

Results - chemical analysis

| Parameter | Value | Unit | Type | Method used | Uncertainty |
|-----------|-------|------|------|-------------------------|-------------|
| Ag | 9,28 | mg/l | A | SOP OV 201 ² | ±20% |

Notice to sampling : The sampling itself is not a subject of accreditation.**Method specification :**

SOP OV 201 (SN EN ISO 17294-1, SN EN ISO 17294-2)

Laboratory workplace :⁽²⁾ - Analyses performed at Ostrava (Partyzánské nám. 7, 702 00 Ostrava)

Methods in TYPE column: "A" accredited test

< - the result is below the detection limit, > - the result is higher than the value presented

Results deal with tested samples only.

Without a written consent of the laboratory, this Report can be reproduced only complete.

These expanded uncertainties of measurement are obtained by multiplying of standard uncertainty of measurement by extending coefficient k=2 (for confidence level 95%). Uncertainty of sampling not included.

Head of Hygienic Laboratories Center : Došká ová Šárka, RNDr.**Checked by :** Lach Karel, Ing. CSc.**Completed by :** Lach Karel, Ing. CSc.**Number of pages :** 1**Date :** 17.10.2016Ing. Vladimíra N mcová
Head of Department of Inorganic Analyses